

Probes for anaesthetic gas scavenging terminal units

IDENTIFICATION

- ▶ Probes for type 1 anaesthetic gas scavenging units (systems with blower or venturi systems)
- ▶ Facility for supply of two different hose connector sizes
- ▶ Aluminium construction with hard anodization surface treatment
- ▶ Laser-engraved production lot and product code



CE MARKING

Notified body: CE 0426
CE marking in risk class II A in accordance with Legislative Decree 24 February 1997, no. 46 "Implementation of Directive 93/42/EEC, concerning Medical Devices" and further modifications



REFERENCE STANDARD

ISO 9170-2: "Terminal units for medical gas pipeline systems - Part 2: Terminal units for anaesthetic gas scavenging systems"

ISO 7396-2: "Medical gas pipeline systems - Part 2: Anaesthetic gas scavenging disposal systems"

ISO 15001: "Anaesthetic and respiratory equipment. Compatibility with oxygen"



* Validity of the standards is referred to the current year








Probes for anaesthetic gas scavenging terminal units

DIMENSIONS



- ▶ Gas: Anaesthetic gas scavenging
- ▶ Storage temperature: -20 °C ÷ +60 °C

- ▶ Working temperature: +10 °C ÷ +40 °C
- ▶ Hose connector for flexible hose
Øi = 22 mm or Øi = 16 mm

CODE	INTERNAL HOSE Ø	NOMINAL PRESSURE bar	PACKAGING		
			 type	 kg	 pcs
PF-06EP-00EV	22 mm	-15 kPa	bag	0.08	1
PF-06EP-10EV	16 mm	-15 kPa	bag	0.08	1

